



Hugh E. Weathers
Commissioner

State of South Carolina Department of Agriculture

Metrology Laboratory

Mailing Address
PO Box 11280
Columbia, SC 29211

Physical Location
237 Catawba Street
Columbia, SC 29201

803.253.4052 (Phone/Fax)
agriculture.sc.gov

CERTIFICATE OF CALIBRATION FOR

2 – 20 kilogram Test Weights

Manufacturer

Rice Lake

Serial Number

See Page 2

State Test Number

SC120703-5-17

Submitted by

Porter Scales
1721 Lake Wheeler Road
Raleigh, NC 27603

The test weights described above have been compared with the standards of the State of South Carolina and were found (adjusted) to be within ASTM Class 2 tolerance with the following exceptions:

Date of Test: 05/23/17

Robert L. McGee, Director
Metrology Laboratory

Comments:

- This calibration was performed within the requirements of the laboratory Quality Management System version 6, Jan 2010 which is compliant with the ISO/IEC 17025, 2005 standard. The laboratory has demonstrated measurement proficiency through training and interlaboratory comparisons and has been issued a NIST Certificate of Measurement Traceability effective through December 31, 2017.
- This calibration has metrological traceability to the International System of Units (SI) through South Carolina primary standards traceable to the National Institute of Standards and Technology through NIST Test Numbers 684/283739-13, 684/288167-16.

Calibration Due Date: 05/31/19

Purchase Order Number: FG04102017

